

CERVICAL SPINE CAGE

Peek material

Supplied sterile

2 anterior and 1 posterior radiographic marker pins

7 heights ranging from 4 mm to 10 mm in 1 mm increments

The cage have serrations on the top and bottom surfaces of the cage

The holes of the cage is intended to hold autogenous bone or bone graft substitute

It is intended for use as an interbody fusion device (from C2-C3 to C7-T1) following anterior cervical discectomy for reduction and stabilization of the cervical spine